

RECEIVED  
CENTRAL FAX CENTER

SEP 14 2009

Application No. 10/562319  
Responsive to the office action dated June 12, 2009

**Amendments to the Claims:**

This listing of claims will replace all prior versions and listings of claims in the application.

**Listing of Claims:**

1. (Currently Amended) An ultrasonic diagnostic apparatus, comprising:
  - a transmission unit that transmits at least one ultrasonic pulse, wherein the at least one ultrasonic pulse is transmitted from a skin surface of a subject toward a blood vessel thereof;
  - a reception unit that receives an ultrasonic echo signal based on the at least one ultrasonic pulse and converts the same into an electric signal, wherein the ultrasonic echo signal is reflected by the blood vessel to obtain a the electric signal of the ultrasonic echo signal representative of a depth direction from the skin surface;
  - a movement detection unit that analyzes a phase of the ultrasonic echo signal, wherein the phase of the ultrasonic echo signal is representative of a direction traversing the blood vessel so that the movement detection unit calculates a movement amount in each of a plurality of parts included in a blood vessel wall constituting the blood vessel and a vicinity of the blood vessel wall; and
  - a boundary detection unit that detects a boundary position, wherein the detected boundary position is between the blood vessel wall and a blood flow region in a lumen of the blood vessel through which blood flows based on a track indicating a variation in the calculated movement amount in each of the plurality of parts and by detecting a position where neighboring parts have tracks with different features.
2. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 1, further comprising a ROI placement unit that sets placement of a ROI, wherein the ROI is set where the boundary position along the depth direction from the skin surface is to be detected by the boundary detection unit,

Application No. 10/562319

Responsive to the office action dated June 12, 2009

wherein the ROI placement unit places the ROI so as to lie over at least one of an anterior wall of the blood vessel wall on a side closer to the transmission unit and a posterior wall of the blood vessel wall on a side farther from the transmission unit.

3. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 1, wherein the transmission unit transmits a plurality of ultrasonic pulses toward a plurality of parts along a longitudinal direction of the blood vessel, and the boundary position detection unit detects the boundary position for each of the plurality of parts along the longitudinal direction of the blood vessel.
4. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 3, further comprising a filter processing unit that performs filter processing, wherein the filter processing unit processes data representing the boundary position along the longitudinal direction of the blood vessel that is detected by the boundary position detection unit.
5. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 3, further comprising a display unit that displays an image, wherein the image is that of the blood vessel in cross section along the longitudinal direction of the blood vessel based on the boundary position along the longitudinal direction of the blood vessel that is detected by the boundary position detection unit.
6. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 1, further comprising an average processing unit that performs average processing of data, wherein the data represents the boundary position that is detected by the boundary position detection unit based on data representing a boundary position obtained a predetermined number of measurement cycles before.
7. (Original) The ultrasonic diagnostic apparatus according to claim 6, wherein the measurement cycles include a heartbeat cycle of a blood flow that flows through the blood vessel.

Application No. 10/562319

Responsive to the office action dated June 12, 2009

8. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 1, further comprising an average processing unit that performs average processing of data, wherein the data represents the movement amount of the blood vessel wall that is detected by the movement detection unit based on data representing a movement amount obtained a predetermined number of measurement cycles before.

9. (Currently Amended) An ultrasonic diagnostic apparatus, comprising:

a transmission unit that transmits at least one ultrasonic pulse, wherein the at least one ultrasonic pulse is transmitted from a skin surface of a subject toward a blood vessel thereof;

a reception unit that receives an ultrasonic echo signal based on the at least one ultrasonic pulse and converts the same into an electric signal, wherein the ultrasonic echo signal is reflected by the blood vessel to obtain the electric signal of the ultrasonic echo signal representative of a depth direction from the skin surface;

a movement detection unit that analyzes a phase of the ultrasonic echo signal, wherein the phase of the ultrasonic echo signal is representative of a direction traversing the blood vessel so that the movement detection unit calculates a movement amount in each of a plurality of parts included in a blood vessel wall constituting the blood vessel and a vicinity of the blood vessel wall; and

a boundary detection unit that detects a boundary position, wherein the boundary position is between an inner membrane of the blood vessel and a blood flow region in a lumen of the blood vessel through which blood flows and a position of a middle membrane of the blood vessel based on a track indicating a variation in the calculated movement amount in each of the plurality of parts and by detecting a position where neighboring parts have tracks with different features.

10. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 9, further comprising a ROI placement unit that sets placement of a ROI, wherein the ROI is set where the boundary position between the inner membrane of the blood vessel and the

Application No. 10/562319

Responsive to the office action dated June 12, 2009

blood flow region and the position of the middle membrane are to be detected along the depth direction from the skin surface by the boundary detection unit,

wherein the ROI placement unit places the ROI so as to lie over at least one of an anterior wall of the blood vessel wall on a side closer to the transmission unit and a posterior wall of the blood vessel wall on a side farther from the transmission unit.

11. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 9, further comprising a calculation unit that measures a thickness, wherein the thickness is measured from the inner membrane to the middle membrane based on the boundary position and the position of the middle membrane.

12. (Original) The ultrasonic diagnostic apparatus according to claim 11, wherein the calculation unit measures the thickness from the inner membrane to the middle membrane based on a variation over time in the boundary position and a variation over time in the position of the middle membrane in one heartbeat cycle.

13. (Original) The ultrasonic diagnostic apparatus according to claim 12, wherein the calculation unit calculates at least one of a maximum value, a minimum value and an average value of the thickness in one heartbeat cycle.

14. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 11, wherein the transmission unit transmits the plurality of the ultrasonic pulses toward the plurality of parts along the longitudinal direction of the blood vessel, and the calculation unit measures the thickness at each of the plurality of parts.

15. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 11, further comprising a display unit that displays a part where a maximum thickness is measured among the thicknesses measured at the plurality of parts.

16. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 9, further comprising an angle correction unit that performs angle correction, wherein the

Application No. 10/562319

Responsive to the office action dated June 12, 2009

angle correction is performed with respect to a value of the thickness corresponding to an angle formed between a measuring direction of the thickness calculated by the calculation unit and a direction perpendicular to the blood vessel wall.

17. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 9, further comprising a stability determination unit that determines stability of the thickness calculated by the calculation unit by comparing the thickness calculated by the calculation unit with a thickness obtained a predetermined number of cycles before.

18. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 9, wherein the transmission unit transmits a plurality of the ultrasonic pulses toward a plurality of parts along a longitudinal direction of the blood vessel, the calculation unit measures the thickness at each of the plurality of parts, and the ultrasonic diagnostic apparatus further comprises a stability determination unit that determines stability of the thickness calculated by the calculation unit by comparing the thicknesses measured at the plurality of parts with each other.

19. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 11, further comprising a unit that displays a value of the thickness calculated by the calculation unit on a monitor.

20. (Previously Presented) The ultrasonic diagnostic apparatus according to claim 9, further comprising a unit that displays the boundary position, wherein the unit displays the boundary position and the position of the middle membrane detected by the boundary detection unit on a monitor.